

# Pengukuran dosis di berbagai organ dan image quality pada pemeriksaan computed tomography (CT) abdomen dengan menggunakan fantom = Measurement of dose in various organs and image quality on computed tomography (CT) abdominal examination using phantom

Siti Nur Hidayati, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20467184&lokasi=lokal>

---

## Abstrak

Penelitian ini menggunakan fantom abdomen in house dengan tujuan mengukur dosis di berbagai daerah organ yaitu hepar, ginjal, reproduksi dan bladder. Pengukuran dosis pada daerah organ dilakukan dengan menggunakan dosimeter gafchromic dan TLD. Selain itu penelitian ini bertujuan untuk mengetahui profil dosis sepanjang sumbu-z dan image quality dengan variasi pitch factor. Faktor eksposi yang digunakan disesuaikan dengan aplikasi klinis abdomen yaitu 130 kV, 80 mAs, rotation 1.5 s dan delay 3 s. Pemilihan parameter pitch factor pada pemeriksaan CT abdomen akan mempengaruhi nilai dosis dan image quality. Variasi Pitch factor yang digunakan 0,8; 1 dan 1.5. Secara umum pengukuran dosis dengan gafchromic dan TLD di berbagai daerah organ memperlihatkan bahwa semakin besar penggunaan pitch factor maka dosis yang didapatkan semakin kecil. Profil dosis sepanjang sumbu-z berbentuk parabola yang simetris dengan dosis maksimum di sekitar 3.45 mGy dan dosis minimum pada awal dan akhir scanning sekitar 3.286 mGy. Hubungan nilai SNR dan slice untuk ketiga nilai pitch 0.8, 1 dan 1.5 pada umumnya sinusoidal dan untuk obyek di daerah kanan dan kiri menunjukkan kurva yang berbeda fase. Demikian juga antara dua titik atas dan bawah. Pengukuran kesesuaian antara citra obyek dengan ukuran obyek sebenarnya dari 512 data diperoleh hasil pada pitch factor 0.8 deviasi diameter 0 – 5 sekitar 50.5 dan selebihnya 49.5 deviasinya diatas 5 . Pada pitch factor 1 deviasi 0 – 5 sekitar 53.5 dan deviasi lebih dari 5 sekitar 46.7 , sedangkan untuk pitch factor 1.5 deviasi 0 – 5 sekitar 68 dan deviasi lebih dari 5 sekitar 32.

.....This study uses in house phantom abdomen with the aim of measuring doses in different regions of the organ namely liver, kidney, reproduction and bladder. Measurement of dose in the organ region is done by using gafchromic and TLD dosimeter. In addition, this study aims to determine the profil dose along the z axis and image quality with variation of pitch factor. The exposure factors were adjusted for the clinical application of abdomen 130 kV, 80 mAs, rotation 1.5 and delay 3 s. Selection of pitch factor parameters on abdominal CT examination will affect the dose value and image quality. Variation of pitch factor used 0.8 1 and 1.5. In general, Measurements dose with gafchromic and TLD in different organ regions showed that the greater the use pitch factor the smaller the dose. Profil doses along the z axis are parabolic shapes symmetrical with maximum doses about 3.450 mGy and minimum doses at the start and end of scanning around 3.286 mGy. The relation of SNR and slice values to the three pitch values 0.8 1 and 1.5 is generally sinusoidal and for the object in the right and left regions show different curves of phase. Likewise between the two points above and below. Measurement of conformity between object image and actual object size from 512 data obtained result on pitch 0.8 deviation of 0 – 5 diameter around 50.5 and 49.5 deviation over 5 . In pitch factor 1 deviation 0 – 5 about 53.5 and deviation more than 5 about 46.7 , while for pitch factor 1.5 deviations 0 – 5 about 68 and deviation more than 5 about 32 .