

Uji diagnostik sifilis menggunakan rapid test standard Q syphilis AB dari spesimen darah kapiler dan serum dibandingkan dengan treponema pallidum hemagglutination assay: Studi pada populasi risiko tinggi = Syphilis diagnostic test using rapid test standard Q syphilis AB of fingerprick whole blood and serum specimens compared to treponema pallidum hemagglutination assay: Study on high risk populations.

Rahma Evasari, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20478569&lokasi=lokal>

---

Abstrak

Sifilis merupakan penyakit infeksi yang disebabkan oleh *Treponema pallidum* subspecies *pallidum* (*T. pallidum*), merupakan penyakit kronis dan bersifat sistemik. Sifilis merupakan penyakit yang progresif dengan gambaran klinis aktif (stadium primer, sekunder, dan tersier) serta periode tidak bergejala (sifilis laten). Sifilis masih merupakan masalah kesehatan di seluruh dunia dengan 80-90% kasus baru terjadi di negara berkembang dengan sedikit atau tidak ada akses diagnostik. Sejumlah besar sifilis tidak bergejala. Akibatnya, sebagian sifilis tidak terdiagnosis dan tidak mendapatkan tatalaksana yang baik, sehingga berpotensi menimbulkan gejala sisa serius, manifestasi sifilis tersier, kardiovaskular, neurologik, oftalmologik, otologik, dan berlanjutnya rantai penularan. Penelitian ini bertujuan untuk menilai kemampuan rapid Test STANDARD Q Syphilis Ab dengan menggunakan spesimen serum dan darah kapiler dibandingkan dengan TPHA dalam mendeteksi sifilis pada populasi risiko tinggi yang terdiri atas waria, lelaki yang berhubungan seksual dengan lelaki, dan wanita penjaja seks di Puskesmas Kecamatan Pasar Rebo, Jakarta Timur. Penelitian ini adalah uji diagnostik dengan dengan rancangan studi potong lintang. Hasil penelitian menggunakan spesimen serum memberikan hasil sensitivitas 91,30%, spesifisitas 97,53%, nilai duga positif 95,45%, nilai duga negatif 95,18%, dan akurasi 95,28% dibandingkan dengan TPHA sebagai baku emas. Hasil pengujian dengan spesimen darah kapiler memberikan hasil sensitivitas 84,78%, spesifisitas 98,77%, nilai duga positif 97,50%, nilai duga negatif 91,95%, dan akurasi 93,70% dibandingkan dengan TPHA sebagai baku emas. Kesesuaian hasil rapid test STANDARD Q Syphilis Ab antara spesimen serum dan darah kapiler sangat baik ( $\kappa = 0,8223$ ). Rapid test STANDARD Q Syphilis Ab dapat dijadikan alternatif uji treponemal dalam menunjang diagnosis sifilis, baik sebagai penapisan rutin maupun konfirmasi hasil uji nontreponemal serta penggunaan spesimen darah kapiler dapat dijadikan alternatif uji treponemal yang lebih cepat dan mudah dilakukan.

<hr>

Syphilis is a chronic and systemic disease caused by *Treponema pallidum* subspecies *pallidum* (*T. pallidum*). Syphilis is a progressive disease with active clinical features (primary, secondary, and tertiary syphilis) and asymptomatic periods (latent syphilis). Syphilis is still a worldwide health problem with 80-90% of new cases occurring in developing countries with little or no diagnostic access. A large number of syphilis are asymptomatic. As a result, some syphilis is undiagnosed and does not get good management, potentially causing serious sequelae, the manifestation of tertiary syphilis, cardiovascular, neurologic, ophthalmologic, otologic, and continuous chain of transmission. This study aimed to assess STANDARD Q Syphilis Ab's rapid test capability using serum and fingerprick whole blood specimens compared with TPHA in detecting syphilis in high-risk populations comprised of transgenders, men who have sex with

men, and female sexual workers in Puskesmas Pasar Rebo. This study is a diagnostic test with a cross sectional study design. The results of this study using serum specimens were sensitivity of 91.30%, specificity of 97.53%, positive predictive value 95.45%, negative predictive value of 95.18%, and accuracy 95.28%, compared to TPHA as the gold standard. Test results with fingerprick whole blood specimens gave sensitivity of 84.78%, specificity of 98.77%, positive predictive value of 97.50%, negative predictive value of 91.95%, and accuracy 93.70%, compared to TPHA as the gold standard. Compatibility of rapid test STANDARD Q Syphilis Ab results between serum and fingerprick whole blood specimens was very good ( $\kappa = 0.8223$ ). Rapid test STANDARD Q Syphilis Ab can be used as an alternative treponemal test in supporting syphilis diagnosis, either as routine screening or confirmation of nontreponemal test result and the use of fingerprick whole blood specimen can be used as treponemal test alternative which is faster and easier to do.