

# Percentage depth dose pada pesawat angiografi = Percentage depth dose in angiography

Reisa Cahaya Putri Wibowo, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20481873&lokasi=lokal>

---

## Abstrak

### <b>ABSTRAK</b><br>

Penelitian ini menggunakan Pesawat Angiografi Siemens Artis Zee untuk mengukur prosentase dosis kedalaman (Percentage Depth Dose, PDD) untuk mempelajari dosis di bawah kulit. Pengukuran PDD dilakukan dengan menggunakan film Gafchromic XR-RV3 yang diletakkan di antara fantom akrilik dengan 6 variasi filter pesawat, 5 variasi tegangan tabung, dan 3 variasi fokus berkas. Hasil penelitian menunjukkan karakteristik dosimetri yang didapatkan bersesuaian dengan teori, dimana titik kedalaman maksimum dan titik persentase dosis <10% semakin dalam dengan meningkatnya tegangan tabung dan filter tambahan, sementara ukuran fokus tidak memberikan pengaruh. Penelitian ini juga menunjukkan bahwa dosis telah diserap sebesar lebih dari 69% oleh tubuh pada kedalaman 150 mm. Disimpulkan juga bahwa film Gafchromic XR-RV3 tidak dapat digunakan dalam pengukuran PDD angiografi dengan posisi permukaan tegak lurus berkas karena faktor buildup. Karenanya, diperlukan studi tambahan untuk menginvestigasi kedalaman buildup pada film Gafchromic XR-RV3 untuk keperluan pengukuran PDD.

<hr>

### <b> ABSTRACT </b><br>

This study used Siemens Artist Zees Angiography to measure the percentage of depth dose (PDD) to investigate dose behaviour under the skin in angiography. The PDD measurements were carried out using the Gafchromic XR-RV3 film positioned between acrylic phantoms with 6 variations of added filtrations, 5 variations in tube voltage, and 3 variations in beam focal spot sizes. The results showed that the dosimetry characteristics obtained were in accordance with the theory, where the maximum depth point and point of <10% dose went deeper with the increase of tube voltage and additional filters, and with the focal spot size having no effect. Results also shown that dose were absorbed by more than 69% by the body at 150 mm depth. It was also concluded that the Gafchromic XR-RV3 film may not be ideal in measuring PDD for angiography with the position of the film perpendicular to the beam, i.e due to horizontal buildup factor. Therefore, additional studies are required to investigate the buildup depth in Gafchromic XR-RV3 film for PDD measurement purposes.