

Toleransi Pengaturan Brightness dan Contrast Periodontitis Apikalis dan Abses Apikalis Dini Pada Radiograf Periapikal-Digital = Brightness and Contrast Tolerance of Apical Periodontitis and Early Apical Abscess Cases in Periapical Digital Radiographs

Mikail, author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20500057&lokasi=lokal>

Abstrak

Latar Belakang: Penerapan teknik image enhancement pada radiografi digital saat ini dilakukan berdasarkan preferensi subjektif pengamat. Pengaturan peningkatan citra dilakukan untuk memperjelas citra radiografi, antara lain dengan mengubah kecerahan dan kontras, sehingga memudahkan interpretasi dalam penanganan kasus. **Tujuan:** Untuk mengetahui toleransi nilai kecerahan dan kontras pada radiografi digital kasus periodontitis apikal dan abses apikal dini yang gambaran radiografinya sulit dibedakan. **Metode:** Menyesuaikan pengaturan peningkatan gambar dengan menyesuaikan nilai kecerahan dan kontras pada nilai +10, +15, -10, -15 yang akan diamati, diproses, dan dianalisis menggunakan perangkat lunak analisis data. **Hasil:** Kisaran nilai yang dapat ditoleransi dalam pengaturan peningkatan dan penurunan kecerahan dan kontras pada kasus periodontitis apikal dan abses apikal dini adalah di bawah +10 dan di bawah -10. **Kesimpulan:** Mengatur kecerahan dan kontras ke nilai di atas 10 untuk peningkatan dan penurunan dapat mengubah informasi diagnostik secara signifikan.

.....**Background:** The application of image enhancement techniques in digital radiography is currently carried out based on the subjective preference of the observer. Image enhancement settings are made to clarify the radiographic image, among others by changing the brightness and contrast, so as to facilitate interpretation in case management. **Objective:** To determine the tolerance for brightness and contrast values on digital radiography of cases of apical periodontitis and early apical abscess whose radiographic features are difficult to distinguish. **Method:** Adjust the image enhancement settings by adjusting the brightness and contrast values at +10, +15, -10, -15 values to be observed, processed, and analyzed using data analysis software. **Results:** The range of tolerable values in the setting of increasing and decreasing brightness and contrast in cases of apical periodontitis and early apical abscess was below +10 and below -10. **Conclusion:** Setting the brightness and contrast to values above 10 for increase and decrease can significantly change the diagnostic information.