

Accuracy of flir one thermal imaging adjunct with clinical evaluation as burn wound depth measurement: clinical observation in senior plastic surgery resident and experienced burn consultant plastic surgeon =  
Akurasi flir one thermal imaging sebagai alat bantu dalam menilai kedalaman luka bakar bersama evaluasi klinis: observasi klinis antara senior residen bedah plastik dan ahli konsultan luka bakar bedah plastik

Tara Sadwika P.J., author

Deskripsi Lengkap: <https://lib.ui.ac.id/detail?id=20505171&lokasi=lokal>

---

## Abstrak

Latar Belakang: Tujuan dari manajemen luka bakar adalah untuk menginisiasi penutupan luka dini atau epitelisasi, dan untuk mencegah komplikasi akibat sepsis. Namun, dari praktik harian kami, diagnosis dini, terutama dalam menentukan kedalaman luka bakar pada fase akut, cukup sulit karena proses luka bakar terus berlangsung. Pengukuran objektif merupakan metode tambahan yang baik untuk membantu dokter mengevaluasi kedalaman luka bakar, misalnya pencitraan termal FLIR ONE. Tujuan penelitian ini adalah untuk mengevaluasi validitas FLIR ONE termografi sebagai alat untuk menilai kedalaman luka bakar, dan keandalan evaluasi klinis dan FLIR ONE yang dilakukan oleh ahli konsultan ahli luka bakar bedah plastik dan senior residen bedah plastik. Metode: Studi diagnostik yang dilakukan dari November 2019 - April 2020 di pusat kami. Dengan kriteria inklusi disebutkan kami melakukan pengamatan dua kali berdasarkan evaluasi klinis dan juga alat bantu FLIR ONE termografi pada luka bakar superfisial dan mid-dermal dalam waktu 48 jam pascalukabakar, dan hari 3-5 pascalukabakar, dengan outcome yaitu evaluasi klinis yang dilakukan oleh ahli bedah plastik konsultan luka bakar berpengalaman di hari ke 7. Data dikumpulkan dan menganalisis validitas dan realibilitas. Hasil: 43 sampel yang diambil dari laki-laki 15 (53,6%) dan perempuan 13 (46,4%), usia rata-rata  $41,82 \pm 13,52$  tahun. Sebagian besar sampel adalah dari wajah 14 (32,6%), dan ekstremitas atas 11 (25,6%). Realibilitas: ICC adalah T1 0,95 dan T3 0,98, menunjukkan angka baik hingga hari ke 7 hari pascalukabakar. Kesenjangan evaluasi klinis antara kedua pengamat (konsultan luka bakar bedah plastic berpengalaman dan residen bedah plastik senior) di T1 adalah 6,9% dan di T3 adalah 9,3%. Tidak ada perbedaan yang signifikan dalam penilaian klinis baik di T1 ( $p = 0,82$ ) dan T3 ( $p = 0,51$ ) dan tidak ada perbedaan yang signifikan dalam pengukuran menggunakan alat FLIR ONE antara dua pengamat baik di T1 ( $p = 0,25$ ) dan T3 ( $p = 0,91$ ). Validitas: AUC dihitung pada T1 adalah 0,72 (95% CI: 0,563 - 0,880)  $p = 0,014$  dengan titik batas T1 pada  $-0,8^\circ C$ , menunjukkan diskriminasi moderat antara kategori penyembuhan yang re-epitelisasi  $\leq 7$  hari dan  $> 7$  hari (sensitivitas 62,5%; spesifisitas 78,9%). Kami menggabungkan evaluasi klinis dan T1 dalam waktu 48 jam setelah luka bakar, penggunaan Flir ONE sebagai alat tambahan meningkatkan sensitivitas menjadi 58,33%, spesifisitas 98% dari evaluasi klinis saja. Probabilitas re-epitelisasi temuan klinis kedalaman luka superficial dengan nilai T1  $> -0,8^\circ C$  memiliki probabilitas tertinggi (90,94%) untuk re-epitelisasi dalam waktu kurang dari sama dengan 7 hari.

Kesimpulan: Penelitian ini menunjukkan validitas dan reliabilitas yang baik dari evaluasi klinis saja dan evaluasi klinis dengan FLIR ONE termografi dalam menilai kedalaman luka bakar. Titik potong kami dalam menentukan kedalaman luka bakar adalah  $-0,8^\circ C$ , dengan hasil probabilitas yang baik untuk membedakan hasil epitelisasi berulang. Penelitian ini juga memberi tahu kami bahwa program residensi bedah plastik di rumah sakit pendidikan kami telah berhasil membangun kompetensi modul yang baik, dan residen memiliki

paparan yang cukup terhadap kasus luka bakar.

.....Background: The aim of the management of burn wound is to initiate early wound closure or epithelization, and to prevent sepsis complication. However, from our daily practice, early diagnosis especially in determining the depth of burn wound in acute phase, is quiet difficult as burn wound process is running. Objective measurement may be great adjunct methods to help clinician evaluating burn wound depth, as an example of FLIR ONE thermal imaging. The objective was to evaluate the validity of FLIR ONE thermal imager as an adjunct tool to assess burn wound depth, and reliability of clinical evaluation and FLIR ONE performed by senior resident of plastic surgery and experienced burn consultant plastic surgeon.

Methods: This is a diagnostic study conducted from November 2019 – April 2020 in our center. With inclusion criteria mentioned we did observation twice based on clinical visual and also FLIR ONE thermal imaging on superficial and mid dermal burn within 48 hours post burn, and post burn day 3-5, outcome by clinical evalution done by experienced burn consultant plastic surgeon on day 7. Data were collected and analyze validity and realibility. Result: We had 43 samples taken from male 15 (53,6%) and female 13 (46.4%), average age  $41.82 \pm 13.52$  years. As facial 14(32.6%), and upper extremities 11 (25.6%) as most samples use.

Reliability: ICCs were T1 0.95 and T3 0.98, indicating excellent reliability up to 7 days after burn. The gap of clinical evaluation between both observers (experienced burn consultant and senior plastic surgery resident) at T1 is 6.9 percent and at T3 is 9.3 percent. There were no significant difference in clinical assessment both in T1 ( $p=0.82$ ) and T3 ( $p=0.51$ ) and no significant difference in measurements using FLIR ONE between two observers both in T1 ( $p=0.25$ ) and T3 ( $p=0.91$ ). Validity: the area under the curve was calculated at T1 was 0.72 (95% CI: 0.563 – 0.880)  $p = 0.014$  with a cut-off point of T1 at  $-0.8^{\circ}\text{C}$ , shows a moderate discrimination between healing categories re-epithelialization  $\leq 7$  days and  $> 7$  days (62.5% sensitivity; 78.9% specificity). We combined clinical evaluation and T1 within 48 hours post burn, the use of Flir One as an adjunct tool increased the sensitivity to 58.33%, specificity 98% of clinical evaluation solely. the probability of re-epithelialization of clinical finding of superficial wound depth with T1 value of  $>-0.8^{\circ}\text{C}$  had the highest probability (90.94%) to re-epithelialized in less equal to 7 days.

Conclusion: This research showed good validity and reliability of clinical evaluation alone and clinical evaluation adjunct with FLIR ONE thermal imaging in assessment of burn wound depth. Our cut off point in determining the burn wound depth was  $-0.8^{\circ}\text{ C}$ , with good probability result to differentiate re-epithelialization outcome. This research told us that plastic surgery residency program of our teaching hospital had successfully established a good module competency, and resident had enough exposure to the burn cases.